

# 593 RELATIVE HUMIDITY SENSOR

## SERVICE MAINTENANCE, REPAIR, SPARE PARTS

Met One Instruments, Inc.; 1600 NW Washington Blvd.; Grants Pass, OR. 97526  
Ph. 541-471-7111; Fax 541-471-7116; E-mail: service@metone.com

### REPAIR AND FIELD MAINTENANCE

A complete repair and service department is maintained at our factory in Grants Pass, Oregon, and at our regional service facility in Dallas, Texas. Most repairs can be made in the field by a simple substitution of components. Some repairs do require recalibration following the replacement of faulty components. If calibration standards are not available, the instrument can be returned to our service facility. The turn around time is generally less than five (5) working days.

### RETURN AUTHORIZATION

Prior to any product return, please contact the factory for a return authorization number.

### BASIC SERVICE(S)

ORDER #	DESCRIPTION	LIST PRICE
SVC0593	593 Relative Humidity Sensor 3 Point Calibration. NIST Traceable Certificate Clean Element and Filter Replace O-Rings as needed	\$ 150.00

**Note: 083C will be upgraded to 083D at time of Service**

### REPLACEABLE PARTS

Note: Replacement of some of the following parts will require re-calibration of the sensor. Those replacement items that require re-calibration of the sensor are indicated with an “\* “. Instructions are provided in the Product User’s Manual.

PART #	DESCRIPTION	LIST PRICE
* 820202	RH SENSOR CHIP, FOR 083D, 083E CAL REQ.	\$ 125.00
560014	FILTER, MEMBRANE	\$ 37.50
860173	NYLON WASHER	\$ 1.00
720050	O RING	\$ 0.50
480510	STRAIN RELIEF	\$ 5.00
1575-5	SKY SHIELD PLATE	\$ 30.00

**\*\*CONSULT FACTORY FOR CABLE PRICING**