

GT-331 AEROSOL MASS MONITOR

SERVICE MAINTENANCE, REPAIR, SPARE PARTS

Met One Instruments, Inc.; 1600 NW Washington Blvd.; Grants Pass, OR. 97526
Ph. 541-471-7111; Fax 541-471-7116; E-mail: service@metone.com

REPAIR AND FIELD MAINTENANCE

A complete repair and service department is maintained at our factory in Grants Pass, Oregon, and at our regional service facility in Dallas, Texas. Most repairs can be made in the field by a simple substitution of components. Some repairs do require recalibration following the replacement of faulty components. If calibration standards are not available, the instrument can be returned to our service facility. The turn around time is generally less than five (5) working days.

RETURN AUTHORIZATION

Prior to any product return, please contact the factory for a return authorization number.

BASIC SERVICE(S)

ORDER #	DESCRIPTION	LIST PRICE
SVC0102	CALIBRATION, FACTORY, GT-331 PARTICLE MONITOR Includes Calibration and NIST Traceable Certificate Inspect Filter, Inspect, Clean Optical Sensor, Pump and Tubing. Cycle and Test Battery; Install any applicable upgrades.	\$ 125.00
SVC3000	AS FOUND TEST & CERTIFICATE - HAND HELD	\$ 50.00

Parts Marked with an Asterisk (*) ahead of the part number will be replaced as needed during the calibration

REPLACEABLE PARTS

PART #	DESCRIPTION	LIST PRICE
8555	RS232 SERIAL CABLE	\$ 36.00
400100	CABLE, AC POWER	\$ 15.00
510640	POWER SUPPLY	\$ 75.00
G3110	ISO-KINETIC SAMPLING PROBE	\$ 55.00
8517	CARRYING CASE	\$ 45.00
993020	MINIATURE SCREW DRIVER	\$ 4.00
GT-331-9800	GT-331 OPERATION MANUAL	\$ 25.00
G3111	ZERO FILTER	\$ 40.00
3306	AIRPLUS SOFTWARE	\$ 50.00
* 8538	IN-LINE FILTER ASSEMBLY	\$ 30.00
* 8548	BATTERY ASSEMBLY	\$ 70.00
* 8650	PUMP ASSY	\$ 250.00