

# GT-526 PARTICLE COUNTER

## SERVICE MAINTENANCE, REPAIR, SPARE PARTS

Met One Instruments, Inc.; 1600 NW Washington Blvd.; Grants Pass, OR. 97526

Ph. 541-471-7111; Fax 541-471-7116; E-mail: service@metone.com

### REPAIR AND FIELD MAINTENANCE

A complete repair and service department is maintained at our factory in Grants Pass, Oregon, and at our regional service facility in Dallas, Texas. Most repairs can be made in the field by a simple substitution of components. Some repairs do require recalibration following the replacement of faulty components. If calibration standards are not available, the instrument can be returned to our service facility. The turn around time is generally less than five (5) working days.

### RETURN AUTHORIZATION

Prior to any product return, please contact the factory for a return authorization number.

### BASIC SERVICE(S)

ORDER #	DESCRIPTION	LIST PRICE
SVC0106	CALIBRATION, FACTORY, GT-526 HAND HELD Includes Calibration and NIST Traceable Certificate Inspect Filter, Inspect, Clean Optical Sensor, Pump and Tubing. Cycle and Test Battery; Install any applicable upgrades.	\$ 200.00
SVC3000	AS FOUND TEST & CERTIFICATE - HAND HELD	\$ 50.00
SVC3002	CALIBRATE G3120 TEMP/RH SENSOR	\$ 60.00

Parts Marked with an Asterisk (\*) ahead of the part number will be replaced as needed during the calibration

### REPLACEABLE PARTS

PART #	DESCRIPTION	LIST PRICE
3228	RS232 SERIAL CABLE	\$ 40.00
400100	CABLE, AC POWER	\$ 15.00
510640	POWER SUPPLY	\$ 75.00
G3110	ISO-KINETIC SAMPLING PROBE	\$ 55.00
8517	CARRYING CASE	\$ 45.00
993020	MINIATURE SCREW DRIVER	\$ 4.00
GT-526-9800	GT-526 OPERATION MANUAL	\$ 25.00
G3111	ZERO FILTER	\$ 40.00
G3120	TEMPERATURE / HUMIDITY SENSOR	\$ 350.00
G3115	PRINTER FOR USE WITH HANDHELD MONITORS	\$ 450.00
750509	PRINTER PAPER	\$ 5.00
* 8538	IN-LINE FILTER ASSEMBLY	\$ 30.00
* 9285	Battery Pack, GT-521, Aerocet 531, GT-526	\$ 90.00
* 8650	PUMP ASSY	\$ 250.00