

# 466A DATALOGGER

## SERVICE MAINTENANCE, REPAIR, SPARE PARTS

Met One Instruments, Inc.; 1600 NW Washington Blvd.; Grants Pass, OR. 97526  
Ph. 541-471-7111; Fax 541-471-7116; E-mail: [service@metone.com](mailto:service@metone.com)

---

### REPAIR AND FIELD MAINTENANCE

A complete repair and service department is maintained at our factory in Grants Pass, Oregon, and at our regional service facility in Dallas, Texas. Most repairs can be made in the field by a simple substitution of components. Some repairs do require recalibration following the replacement of faulty components. If calibration standards are not available, the instrument can be returned to our service facility. The turn around time is generally less than five (5) working days.

### RETURN AUTHORIZATION

Prior to any product return, please contact the factory for a return authorization number.

### BASIC SERVICE(S)

ORDER #	DESCRIPTION
SVC0466	DATALOGGER SERVICE, 466A  <i>Includes: Test for proper function, program sensor channels as required, labor to replace parts as needed.</i>

### REPLACEABLE PARTS

PART #	DESCRIPTION
10193	AUTOMET 466A LOGGER COMPLETE FRONT PANEL ASSY <i>Includes: Assembled Front Panel, Display, Keypad, with Overlay</i>
3230-7	PCBA, CPU,AUTOMET
3240-3	PCBA, SENSOR INTERFACE
9438-1	POWER SUPPLY, 466A, 15V/2A, WEATHERPROOF ENCL.

**\*\*CONSULT FACTORY FOR CABLE PRICING**