

# GT-526 PARTICLE COUNTER

## SERVICE MAINTENANCE, REPAIR, SPARE PARTS

Met One Instruments, Inc.; 1600 NW Washington Blvd.; Grants Pass, OR. 97526  
Ph. 541-471-7111; Fax 541-471-7116; E-mail: service@metone.com

---

### REPAIR AND FIELD MAINTENANCE

This instrument must be returned to the factory for Calibration. The turn around time is generally less than five (5) working days.

### RETURN AUTHORIZATION

Prior to any product return, please contact the factory for a return authorization number.

### BASIC SERVICE(S)

ORDER #	DESCRIPTION	LIST PRICE
SVC0106	CALIBRATION, GT-526 PARTICLE COUNTER Includes Calibration and NIST Traceable Certificate Inspect Filter, Inspect, Clean Optical Sensor, Pump and Tubing. Cycle and Test Battery; Install any applicable upgrades. <i>Includes "As Received" and "As Left" NIST Cal certificates</i>	\$ 275.00
SVC3002	BENCH CAL, G3120 TEMP/RH SENSOR, NIST	\$ 65.00
SVC0109	ISO 21501-4 CALIBRATION	\$ 650.00

Parts Marked with an Asterisk (\*) ahead of the part number will be replaced as needed during the calibration

### REPLACEABLE PARTS

PART #	DESCRIPTION	LIST PRICE
3228	RS232 SERIAL CABLE	\$ 40.00
400100	CABLE, AC POWER	\$ 15.00
510640	POWER SUPPLY	\$ 75.00
G3110	ISO-KINETIC SAMPLING PROBE	\$ 55.00
8517	CARRYING CASE	\$ 45.00
993020	MINIATURE SCREW DRIVER	\$ 4.00
GT-526-9800	GT-526 OPERATION MANUAL	\$ 25.00
G3111	ZERO FILTER	\$ 45.00
G3120	TEMPERATURE / HUMIDITY SENSOR	\$ 350.00
G3115	PRINTER FOR USE WITH HANDHELD MONITORS	\$ 450.00
750509	PRINTER PAPER	\$ 3.50
* 8538	IN-LINE FILTER ASSEMBLY	\$ 30.00
* 9285	Battery Pack, GT-521, Aerocet 531, GT-526	\$ 90.00
* 8650	PUMP ASSY	\$ 250.00