

GT-521 PARTICLE COUNTER

SERVICE MAINTENANCE, REPAIR, SPARE PARTS

Met One Instruments, Inc.; 1600 NW Washington Blvd.; Grants Pass, OR. 97526
Ph. 541-471-7111; Fax 541-471-7116; E-mail: service@metone.com

REPAIR AND FIELD MAINTENANCE

This instrument must be returned to the factory for Calibration. The turn around time is generally less than five (5) working days.

RETURN AUTHORIZATION

Prior to any product return, please contact the factory for a return authorization number.

BASIC SERVICE(S)

ORDER #	DESCRIPTION
SVC0103	CALIBRATION, GT-521, GT-521S HAND HELD Includes Calibration and NIST Traceable Certificate Inspect Filter, Inspect, Clean Optical Sensor, Pump and Tubing. Cycle and Test Battery; Install any applicable upgrades. <i>Includes "As Received" and "As Left" NIST Cal certificates</i>
SVC3002	BENCH CAL, G3120 TEMP/RH SENSOR, NIST
SVC0109	ISO 21501-4 CALIBRATION

Parts Marked with an Asterisk (*) ahead of the part number will be replaced as needed during the calibration

REPLACEABLE PARTS

PART #	DESCRIPTION
3228	RS232 SERIAL CABLE
400100	CABLE, AC POWER
510640	POWER SUPPLY
G3110	ISO-KINETIC SAMPLING PROBE
8517	CARRYING CASE
993020	MINIATURE SCREW DRIVER
3305	GT-COMM SOFTWARE
GT-COM-9800	GT-COMM OPERATION MANUAL
GT-521-9800	GT-521 OPERATION MANUAL
G3111	ZERO FILTER
G3120	TEMPERATURE / HUMIDITY SENSOR
G3115	PRINTER FOR USE WITH HANDHELD MONITORS
750509	PRINTER PAPER
* 8538	IN-LINE FILTER ASSEMBLY
* 9285	Battery Pack, GT-521, Aerocet 531, GT-526
* 8650	PUMP ASSY