

BT-610 BENCH TOP PARTICLE MONITOR

SERVICE MAINTENANCE, REPAIR, SPARE PARTS

Met One Instruments, Inc.; 1600 NW Washington Blvd.; Grants Pass, OR. 97526

Ph. 541-471-7111; Fax 541-471-7116; E-mail: service@metone.com

REPAIR AND FIELD MAINTENANCE

This instrument must be returned to the factory for Calibration. The turn around time is generally less than five (5) working days.

RETURN AUTHORIZATION

Prior to any product return, please contact the factory for a return authorization number.

BASIC SERVICE(S)

| ORDER # | DESCRIPTION |
|---------|---|
| SVC0108 | CALIBRATION, BT-610, BT-637, BT-637S BENCH-TOP Includes Calibration and NIST Traceable Certificate Inspect Filter, Inspect, Clean Optical Sensor, Pump and Tubing. Cycle and Test Battery; Install any applicable upgrades. <i>Includes "As Received" and "As Left" NIST Cal certificates</i> |
| SVC0109 | ISO 21501-4 CALIBRATION |

Parts Marked with an Asterisk (*) ahead of the part number will be replaced as needed during the calibration

REPLACEABLE PARTS

| PART # | DESCRIPTION |
|--------|--|
| 81643 | PUMP MODULE (BT-610, BT-645) |
| 500784 | USB CABLE |
| 550065 | RS232 CABLE, VARIOUS PRODUCTS. |
| 80642 | BATTERY PACK, BT-610, BT-637, BT-637S |
| 80643 | BATTERY CHARGER, BT-610, BT-637, BT-637S |
| G3110 | ISO-KINETIC SAMPLING PROBE |
| G3111 | ZERO FILTER |
| G3120 | TEMPERATURE / HUMIDITY SENSOR |
| 750514 | THERMAL PRINTER PAPER |
| 9801 | FLOW METER |